



Fluoroscopic Devices (Art. 18.2 of SI 478 (2002))

The Council noted satisfactory correspondence from the Faculty of Radiologists, (SR/DS 5/9/03), confirming that no fluoroscopy equipment without image intensification is in use within the State and that there are no circumstances that would warrant the use of fluoroscopic devices which do not have a dose control device.

(Adopted by Medical Council 3 September 2004)